



CERTIFIED WEIGHT REPORT

产品编号 **Part Number:** 71412
产品批号 Lot Number: 031824
Description: Tetraethyllead

Solvent(s): **Lot#**
Methanol EH299-US

		031824
Formulated By:	Lawrence Barry	DATE
		031824
Reviewed By:	Pedro L. Rentas	DATE

(保质期: 月月日日年年) Expiration Date: 031829
Recommended Storage: Refrigerate (4 °C) (推荐保存条件: 4)
Nominal Concentration (µg/mL): 1000
NIST Test ID#: 6UTB


Weight(s) shown below were combined and diluted to (mL): 200.0 0.002 Balance Uncertainty
5E-05 Flask Uncertainty

Expanded **SDS Information**
(Solvent Safety Info. On Attached pg.)



Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
(实际浓度)(扩展不确定度)												
1. Tetraethyllead	1412	15309DO	1000	99.99	0.2	0.20000	0.20018	1000.9	4.0	78-00-2	0.075mg/m3/8H(skin)	ori-rat 12300ug/kg

Method GC6MSD-1: Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. Analysis performed by Candice Warren.

TIC: [BSB3]71412.D



Scan 3155 (48.054 min): [BSB3]71412.D



- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).